

Performance Evaluation of the INSTI HIV-1/2 Antibody Point-of-Care Test in Early and Established Infections

Sarah Adams B.S.

Biologist ORISE Fellow

2016 HIV Diagnostics Conference

3/23/2016

Why Rapid Tests?

- Affordable
- Easy to use
- Accessible testing
- Fast results



- Increase awareness of HIV status
- Provide opportunities for faster linkage to care and treatment
- Minimize transmission

BUT they MUST be accurate!

Low sensitivity of a Rapid Test (RT) during the early stages of HIV infection **increases the risk of misdiagnosing** patients with active viral replication and high risk transmission

INSTI™ HIV-1/2 Antibody Test (INSTI)

- Approved in the U.S. for finger stick (CLIA waived) and venous whole blood and plasma for HIV-1 and HIV-2 antibody testing
- Flow-through HIV-1/HIV-2 Rapid Antibody test
- Test time: 60 seconds



Performance characteristics

Sample Type	Sensitivity	Specificity
Venous Whole Blood	99.9%	100%
Plasma	99.9%	100%
Fingerstick Blood	99.8%	99.5%
CLIA Fingerstick Blood*	100%	99.8%

- Recent study showed that sensitivity of a rapid point of care assay for early HIV antibody detection is enhanced by its affinity for HIV gp41 IgM antibodies**

* Galli RA, Green KF, La Marca A, et al.. J Clin Virol 2013;58 Suppl 1:e65-9.

** Moshgabadi N., Galli R., Daly A., Ko S.M.S., Westgard T., Bulpitt A., Shackleton C.R, et al J Clin Virol 2015

INSTI

- Successfully evaluated in several studies outside of the United States
 - Performed well:
 - compared to other POC tests for the detection of early HIV infection using residual serum
 - at detecting HIV exposure in children younger than 18 months and HIV diagnosis in children older than 18 months with whole blood



Objective

- To evaluate the performance of INSTI in a laboratory using plasma and/or simulated whole blood specimens from early and established HIV infections

Materials and methods


- Plasma from commercial HIV-1 seroconversion panels
 - Estimated relative sensitivity of INSTI and other assays using a 50% cumulative frequency analysis with Western blot positive (n=166)
 - Paired comparison of the performance of INSTI with other FDA-approved RTs using McNemar's statistical test (n=230)
- Eight HIV-1 seroconversion plasma panels (n=40) were used to simulate whole blood
 - INSTI results were compared to results generated with two Ag/Ab Combo tests and one IgG/IgM immunoassay

Methods

- Plasma from HIV-1 positive individuals at different stages of HIV-1 infection
 - Screening Targeted Populations to Interrupt On-going Chains of HIV Transmission with Enhanced Partner Notification (STOP) Study
 - Specimens were tested with INSTI at CDC and compared to Abbott ARCHITECT, Bio-Rad Multispot, Abbott HIV-1 RNA viral load results
- Plasma specimens from
 - Individuals with established HIV-1 and HIV-2 infections (western blot positive) were used to calculate sensitivity
 - HIV-negative individuals were used to calculate specificity
- Plasma specimens from HIV-1 positive individuals who receive antiretroviral therapy (ART)

RESULTS

Assay reactivity during early HIV-1 infections using 50% cumulative frequency analysis



- NAAT
- Ab EIA
- Ab Lateral flow
- Ag/Ab tests
- Ab flow through
- Ab Dual Path Platform


Paired comparison analysis of INSTI and other FDA-approved RTs

<u>Test 1</u>	Number of results test 1/results INSTI				<u>p value</u>
	<u>NR/NR</u>	<u>NR/R</u>	<u>R/NR</u>	<u>R/R</u>	
Determine Ag/Ab	102	2	34	92	<0.0001*
Multispot ^a	133	6	2	88	0.289
Reveal G2/G3 ^a	134	14	1	80	0.0019*
DPP HIV-1/2	134	15	2	79	0.0036*
Statpak	134	20	1	74	<0.0001*
Complete	135	22	1	74	<0.0001*
Unigold	136	23	0	70	<0.0001*
Oraquick Advance	134	34	2	60	<0.0001*

^a 229 total samples available for analysis; R: reactive, NR: not reactive

* statistically significant

Performance with 37 HIV-1 Simulated Whole Blood and Plasma Specimens from 8 Seroconverters



- **INSTI in plasma and simulated whole blood**
 - No delay in early HIV-1 infections
 - 3 samples showed discrepant results

Performance of INSTI with Simulated Whole Blood and Plasma from HIV-1 Seroconverters

- False reactive with whole blood

ID #	Days between bleeds	NAAT	IA	IA	RT		INSTI	
			Ag/Ab	IgG/IgM	Ag	Ab	plasma	whole blood
924-04		nonreactive	-	-	-	-	-	+
924-05	16	reactive	+	-	+	-	-	-
924-06	7	reactive	+	+	+	+	+	+
924-07	2	reactive	+	+	-	+	+	+
924-08	5	reactive	+	+	-	+	+	+

ID #	Days between bleeds	NAAT	IA	IA	RT		INSTI	
			Ag/Ab	IgG/IgM	Ag	Ab	plasma	whole blood
931-05		reactive	+	-	-	-	-	+
931-06	13	reactive	+	+	+	+	+	+
931-07	5	reactive	+	+	-	+	+	+
931-08	2	reactive	+	+	-	+	+	+
931-09	7	reactive	+	+	-	+	+	+
940-01		reactive	+	-	-	-	-	-
940-02	7	reactive	+	-	+	-	-	+
940-03	4	reactive	+	+	+	+	+	+
940-04	4	reactive	+	+	-	+	+	+
940-05	3	reactive	+	+	-	+	+	+
944-01		reactive	-	-	-	-	-	-
944-02	2	reactive	-	-	-	-	-	-
944-03	5	reactive	+	-	+	-	-	+
944-04	2	reactive	+	-	+	-	-	+
944-05	5	nonreactive	+	+	-	+	+	+
944-06	2	reactive	+	+	-	+	+	+

INSTI Reactivity in Early infection

NAT Positive/ ARC Positive

Plasma

Multispot

StatPak

Plasma

Whole blood

Total

Positive

Negative

-

-

51

7

44

-

+

2

0

2

-

NA

21

8

13

ind

-

9

6

3

ind

+

1

1

0

ind

NA

1

1

0

- INSTI with plasma detected more early HIV-1 infections than Statpak with fingerstick whole blood

INSTI Reactivity in Established infection

<u>ARC Positive</u> plasma		INSTI Results		
Multispot plasma	StatPak Whole blood	<u>Total</u>	<u>Positive</u>	<u>Negative</u>
+	-	9	9	0
+	+	174	174	0
+	NA	20	20	0

- INSTI detected all established infections including 9 that were StatPak negative with fingerstick whole blood

Sensitivity of INSTI HIV-1/HIV-2 antibody test

	INSTI results				Sensitivity (%)	95% Confidence Intervals
	Total	Reactive	Non- reactive	Repeatedly Invalid		
HIV-1 Group M						
HIV-1 B subtype	501	500	1	0	99.80	98.88- 99.96
HIV-1 non-B subtype	111	110	0	1	100	96.63- 100
HIV-1 Group O	3	3	0	0	100	43.85- 100
HIV-2	80	80	0	0	100	95.42- 100

- Overall sensitivity **99.84%** [99.08- 99.97% 95%CI]
- Specificity **99.80%** [98.87- 99.96% 95% CI]

Plasma specimens from HIV-1 positive individuals who receive ART

- 35 HIV-1 positive donors with HIV-1 Viral load results (Abbott m2000)
 - 21 Target not detected
 - 5 <40 copies/ml
 - 9 mean of 163 copies/ml (range:48- 391 copies/ml)
- All of these plasma specimens were reactive with INSTI

Summary

- In HIV-1 early infections, INSTI was positive 9 days before a positive WB
- Flow-through RTs detected significantly more infections than the antibody-based lateral flow FDA-approved RTs
- Using simulated whole blood, INSTI performed similarly to plasma with no delay observed between plasma and whole blood among early HIV-1 infection
- Among cross-sectional plasma samples, INSTI detected 23 (27%) of 85 specimens from early HIV-1 infections and 203 (100%) specimen from established infections
- The sensitivity was 99.84% for HIV-1 and 100% for HIV-2 and the specificity was 99.80% in plasma specimens
- All 35 HIV-1 positive plasma donors who receive ART were reactive with INSTI

Conclusions

- INSTI performed significantly better than antibody-based lateral flow RTs at detecting early HIV-1 infections (IgM detection), and performed slightly better than the other flow through RTs
- Sensitivity and specificity were within the manufacturer's reported ranges
- This study was limited to simulated whole blood. Further evaluation with fingerstick/venipuncture whole blood is needed to evaluate point of care performance
- The ease of use of INSTI and the almost instant result makes the test a good option for point-of-care settings

Acknowledgements

- Wei Luo
 - Laura Wesolowski
 - Philip Peters
 - S. Michele Owen
 - Silvina Masciotra
- STOP authors:
- Emily Westheimer
 - Cindy Gay
 - Lisa B. Hightow-Weidman
 - Stephanie Cohen

Thank you 😊

For more information please contact Centers for Disease Control and Prevention

1600 Clifton Road NE, Atlanta, GA 30333

Telephone: 1-800-CDC-INFO (232-4636)/TTY: 1-888-232-6348

Visit: www.cdc.gov | Contact CDC at: 1-800-CDC-INFO or www.cdc.gov/info

The findings and conclusions in this report are those of the authors and do not necessarily represent the official position of the Centers for Disease Control and Prevention.